



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Savant Labs

4800 James Savage Road, Midland, MI 48642

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

Chemical and Mechanical Testing *(As detailed in the supplement)*

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Initial Accreditation Date:

Issue Date:

Expiration Date:

December 11, 2015

January 2, 2018

March 31, 2020

Accreditation No.:

Certificate No.:

84229

L18-1

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjllabs.com



Certificate of Accreditation: Supplement

Savant Labs

4800 James Savage Road, Midland, MI 48642
 Contact Name: Maggie Smerdon Phone: 989-496-2301

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Chemical ^F	Engine Oils and Lubricants	Kinematic	ASTM D445 ASTM D2270	0.2 cSt to 300 000 cSt
		Dynamic Viscosities	ASTM D4683 ASTM D6616	1 cP to 25 cP
			ASTM D5133	1 500 cP to 100 000 cP
			ASTM D2983	300 cP to 900 000 cP
			ASTM D5293	900 cP to 25 000 cP
		Sulfur Analysis by UV Fluorescence	ASTM D5453	1 ppm to 8 000 ppm
		Inductively Coupled Plasma Atomic Emission Spectroscopy	ASTM D5185 ASTM D4951	5 ppm to 10 000 ppm
		Boiling Point Distribution and Estimation of Engine Oil Volatility by Gas Chromatography	ASTM D6417 ASTM D2887EXT	Volatility 1.8 % to 19.8 %
		Chlorine Content by XRF	ASTM D6443	5 mg/kg to 250 mg/kg
		FTIR Oxidation, Soot (Phosphate and Sulfate), and Nitration	E2412: D7414, D7415, D7412, D7844, D7624	Differential Trend Analysis with Reference
		Nitrogen by Chemiluminescence	ASTM D5762 & ASTM D4629	1 ppm to 10 000 ppm
		Elastomer properties	ASTM D7216 CEC L-112	100 °C to 150 °C
		Base Number	ASTM D2896 ASTM D4739	0.1 mg KOH/g to 250 mg KOH/g
		Water by Karl Fischer	ASTM D6304	10 mg/kg to 25 000 mg/kg
		Acid Number	ASTM D664	0.1 mg KOH/g to 150 mg KOH/g
Mechanical ^F		Flash and Fire Point	ASTM D92	79 °C to 400 °C
		Flash Point	ASTM D93	40 °C to 370 °C
		Foam Sequence I-III	ASTM D892	0 mL to 800 mL
		Foam Sequence IV	ASTM D6082	0 mL to 800 mL
		Pour Point	ASTM D97	20 °C to -60 °C
		Sulfated Ash	ASTM D874	0.05 % to 25 %
		Evaporation Loss of Lubricating Oils	ASTM D5800	0 % to 25 %
		Density	SAVLAB Density by Pycnometer	Temperatures of -70 °C to 150 °C



Certificate of Accreditation: Supplement

Savant Labs

4800 James Savage Road, Midland, MI 48642
Contact Name: Maggie Smerdon Phone: 989-496-2301

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Mechanical ^F	Engine Oils and Lubricants	Kurt Orbahn 30 Pass 90 Pass	ASTM D6278 ASTM D7109	Pumpable fluids < 30 cP @ 100 °C
		Dynamic Viscosities	ASTM D4684 ASTM D3829	5 000 cP to 400 000 cP
		Fuel Dilution	ASTM D3525 ASTM D3524	0.1 % to 100 %

1. The presence of a superscript F means that the laboratory performs testing of the indicated parameter at its fixed location. Example: Outside Micrometer^F would mean that the laboratory performs this testing at its fixed location.

